



**CORRIGENDUM**

With reference to tender no: AIIMS/Pat/Tender/2016/PaediaticSur/Urethroscope/346 dated 19/01/2017, regarding Tender for Supply & Installation of Urological Equipments for Department of Paediatic Surgery at AIIMS Patna, a pre-bid meeting was held on 27/01/2017 in committee hall, medical college building, AIIMS Patna.

The issue raised by the firms were discussed and the committee recommended for the following corrigendum to be issued:-

1. Item no. 1 to 30 deleted & following is added
2. The system should have integration of both ultrasonic and / or high frequency bursts of mechanical wave energies delivered simultaneous from a single probe with shock pulse technology & suction control.
3. Generator specifications: Voltage of 90-264 VAC, Frequency-50/60 Hz
4. The system must have lithotripsy probe diameter starting from 0.8mm to 4mm & length should be 300mm-600mm long.
5. The system must offer a complete probe size portfolio with single use & reusable option
6. The system should be compatible with ureterorenoscope & renoscope.
7. The system should be supplied with compatible & standard suction tubes.
8. Sterilization box for lithotripsy hand pieces.
9. The system must have options for using foot switch in addition to hand control/hand switch.
10. Stone catcher set should be provided with the system.
11. System should be US FDA certified
12. *The last date of submission of tender documents is extended up to 17-03-2017, 14.00 hrs and it will open on 17/03/2017 at 15:30 hrs.*

sd/=

**Faculty In-charge  
Procurement Cell  
AIIMS, Patna**