



CORRIGENDUM

With reference to tender no: AIIMS/Pat/Ophthal/Instrument/2016-17/1/409 dated 22/04/2017, regarding Supply & Installation of Non Contact Tonometer For Department of Ophthalmology, a pre-bid meeting was held on 22/04/2017 in committee hall, medical college building, AIIMS Patna.

The issue raised by the firms were discussed and the committee recommended for the following corrigendum to be issued:-

1. The following changes in specification should be as:
 - a. **Point no. 03:** TV monitor screen **should be read as:** Inbuilt LCD Monitor
 - b. **Point no. 04:** R/L change over: automatically detected and displayed **should be read as:** R/L change over: manually or automatically detected and displayed.
2. *The last date of submission of tender documents is extended up to 25-05-2017, 14.00 hrs and it will open on 25/05/2017 at 15:30 hrs.*

sd/=

**Faculty In-charge
Procurement Cell
AIIMS, Patna**