



**CORRIGENDUM**

With reference to tender no: AIIMS/Pat/Tender/PC/2015/RadioDiag/CRSystem/290 dated 3/3/2017, regarding Tender for Supply & Installation of Computed Radiography (CR) System for Department of Radiodiagnosis at AIIMS Patna, a pre-bid meeting was held on 8/3/2017 in committee hall, medical college building, AIIMS Patna.

The issue raised by the firms were discussed and the committee recommended for the following corrigendum to be issued:-

1. Under technical specification of equipment
  - a. Point 10, Line 2 "I long view cassette 120 cm length or more" should be read as "I long view cassette 80 cm length or more with automatic and manual image stitching software for producing total image length of 120 cm or more."
  
2. *The last date of submission of tender documents is extended up to 11-04-2017, 14.00 hrs and it will open on 11/04/2017 at 15:30 hrs.*

sd/=

**Faculty In-charge  
Procurement Cell  
AIIMS, Patna**