



**CORRIGENDUM**

With reference to tender no: AIIMS/Pat/Tender/PC/2015/Physiology/CASA/300 dated 3/3/2017, regarding Tender for Supply & Installation of Computer Assisted Semen Analysis System for Department of Physiology, a pre-bid meeting was held on 8/3/2017 in committee hall, medical college building, AIIMS Patna.

The issue raised by the firms were discussed and the committee recommended for the following corrigendum to be issued:-

1. Under technical specification of equipment
  - a. To rationalize the cost factor, The Term 'fully motorized stage' should be replaced with full automation using manual stage.
  - b. 'Manual counter of any biological sample 'should be replaced with 'micrometer attachment for eye piece for manual counting' and optionally Meckler counter may also be quoted.
  - c. Under microscope head, 'with powerful LED illumination' should be added.
  - d. The approved brands for camera should include 'POINT GREY'.
  - e. In microscope head the specification of microscope may be modified as-Trinocular positive phase contrast microscope with an optional upgradation to fluorescent one.
  
2. *The last date of submission of tender documents is extended up to 04-05-2017, 14.00 hrs and it will open on 4/05/2017 at 15:30 hrs.*

sd/=

**Faculty In-charge  
Procurement Cell  
AIIMS, Patna**