



22/09/2017

## Corrigendum

With reference to tender no: AIIMS/Pat/Tender/Radiology/X-Ray/F4883 dated 3/9/2017, regarding Supply & Installation of High Frequency Mobile X-Ray Machine for Department of Radiodiagnosis, a pre-bid meeting was held on 11/9/2017 in committee hall, medical college building, AIIMS Patna.

The issue raised by the firms were discussed and the committee recommends for the following corrigendum to be issued:-

The following changes in specification should be made:

1. **Page no 37, point 1.** “Xray Generator: 60 mA –100 KVP, High Frequency Xray Generator – 2.5 KW or more” **should be read as** “Xray Generator: 100 mA –100 KVP, High Frequency Xray Generator – 4 KW or more”
2. **Page no 37, point 2.** “Frequency: 40 KHZ or more” **should be read as** “Frequency: 200 KHZ or more”
3. **Page no 37, point 3.** “mAs: upto 200 mAs” **should be read as** “mAs: 200 mAs or more”
4. **Page no 37, point 5.** “Exposure Time: 5 ms to 2.5 seconds or better” **should be read as** “Exposure Time: 10 ms to 5 seconds or better”
5. **Page no 37, point 12C.** “Lead lined storage box to be provided – suitable for minimum 8 cassettes” **should be read as** “Lead lined storage box to be provided – suitable for minimum 6 cassettes ”
6. **Page no 38, point 16a.** “The firm should be ISO certified and providing CE / FDA certified equipment” **should be read as** “The firm should be ISO certified and providing AERB certified equipment ”

7. *The last date of submission of tender documents is extended up to 09-10-2017, 14.00 hrs and it will open on 09/10/2017 at 15:30 hrs.*

sd/=

Faculty In-charge  
Procurement Cell  
AIIMS, Patna