



## CORRIGENDUM

Date: 28/10/2016

**With reference to Tender No. AIIMS/Pat/Tender/ Radiotherapy/USG/592, dated 08/10/2016, a pre bid meeting was held on 17/10/2016 at 15:00 Hrs. in Committee Room Main Building AIIMS Patna.**

The issue raised by the firm where discussed and the committee recommended for the following corrigendum:

1. In Point no. - 4, should be read as Transducers: Five 1) Convex 5 -2 MHZ. for abdominal imaging, 2) Linear 5-10 +/- 1 MHZ for small part and vascular imaging .3) Endocavitary 8-5 MHZ for transvaginal and transrectal ultrasonography and end firing biopsy –one each. 4) Hockey stick transducer 15-6 MHZ for superficial structures. 5) Cardiac probe of 2-5 MHZ.
2. In point no 7, It should be read as: Imaging modes of real time 2D, M- mode colour Doppler, Pulsed wave Doppler, continuous wave Doppler, Power (energy) Doppler and duplex/triplex Doppler should be available.
3. In point no 15- , should be read as: Flat LCD/LED control panel of 10 inches or more.
4. In point no 20 -: should be read as: Facility for storage on CDR /DVD/Pendrive should be available. Data should be transferable through the network to any other workstation.
5. In point no. - 23, it should be read as: The machine should be DICOM ready and should be possible to connect to the Hospital network through suitable LAN port. It will be the responsibility of the supplier to ensure hassle free networking as and when requested.
6. *The last date of submission of tender documents is extended up to 18-11-2016 up to 14.00 hrs*

sd/-

**Faculty In-charge  
Procurement Cell  
AIIMS, Patna**