



Date: 05/12/2016

Corrigendum/Addendum

With reference to Tender No. AIIMS/Pat/Pharma/Instruments/356/IV/2016/708 dated 01/08/2016 a pre bid meeting was held on 21.11.2016 in committee hall, medical college building, AIIMS Patna. A Corrigendum was issued on 29.11.2016 at 3:30pm.

The issues raised by the firms after this corrigendum were discussed and the committee recommended further following corrigendum/Addendum to be issued:-

1. For every item mentioned in schedule of requirement on page 26 of Tender document compliance Report to be submitted in a tabulated and point wise manner clearly mentioning and highlighted the page/para number of original catalogue/data sheet. Any point, if not substantiated with authenticated catalogue/manual, will not be considered
2. In technical specifications - Item No. 8, Non Invasive BP Monitor
 - a) Biopac Non Invasive BP Monitor should be read as Non Invasive BP Monitor
 - b) Addendum as point M: Blood pressure monitor with restrainer, photo electric sensor, and built-in warmer for rat/mice.
 - c) Addendum as point N: blood pressure system should come with complete data acquisition software and data transfer through USB.
3. In technical specifications - Item No. 7, Plethysmometer to be read as follows -
 - a) Plethysmometer should have a resolution of 0.01 ml
 - b) Should be suitable for both rats as well as mice
 - c) Unit should be made from clear acrylic material which should allow obstruction-free viewing
 - d) It should have digital display of the paw volume
 - e) Should allow hands free operation including foot switch control.
 - f) Software supplied with the unit should be able to automatically record results on a PC through a USB/Serial port.
 - g) The system should include as standard a volume transducer with its related calibrators and wetting solution (if needed). Supplied with suitable UPS with 30 minute backup.
 - h) Dust protection cover for control unit.
 - i) Instrument should be CE compliant.
4. The last date of submission of tender documents is extended up to 15.12.2016 up to 14:00 hours.

sd/
Faculty In-charge
Procurement Cell
AIIMS, Patna